

ARCS Instrument Replacement Form

Fill out the following form when replacing instruments at an ARCS site:

System: Skyrad NIP

Site: Manus

OLD Configuration

Part Number	Serial Number	Location	Configuration Date	Cal Gain Coefficient (Responsivity at 450)X10-6	Offset	Date of change UTC	Time Start UTC	Time End UTC
NIP	31346E6	Skyrad	2 Aug 2004	8.1573		3 Nov 2004	23:00	00:20 4 Nov 2004

NEW Configuration

Part Number	Serial Number	Location	Configuration Date	Cal Gain Coefficient (Responsivity at 450)X10-6	Offset	Copy Config. Logger to PC
NIP	29868E6	Skyrad	041104	8.0380		yes

Replacement completed by: Troy Culgan & Rex Pearson

NOTES:

Old Part remains on site as spare: ☐

Old part packed for shipment to: SGP

Other old part disposition (explain):

Other NOTES:
FM(OPS)-008.001